



# EC-type examination certificate

Number **T10150** Revision 1  
Project number: 805513  
Page 1 of 1

Issued by NMI Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands

Notified Body number 0122

In accordance with The "Metrologiewet" (Stb. 2006, 137) as Dutch implementation of Directive 2004/22/EC on measuring instruments (MID).

Manufacturer Kalekalip Makina ve Kalip Sanayi A.S.  
Sefaköy, Tefvikbey Mah. Istiklal Cad. No:13  
34295-K. Çekmece Istanbul  
Turkey

In respect of **A diaphragm gas meter**  
Manufacturer : Kalekalip Makina ve Kalip Sanayi A.S.  
Type : SI-G1,6, SI-G2,5 and SI-G4

Characteristics Destined for the measurement of : gas volume  
Accuracy class : Class 1,5  
Temperature range : -25°C / +55°C  
Mechanical environment class : M2  
Electromagnetic environment class : E1

G-value	Q <sub>max</sub> [m <sup>3</sup> /h]	Q <sub>min</sub> [m <sup>3</sup> /h]	Q <sub>t</sub> [m <sup>3</sup> /h]
4	6	0,040	0,6
2,5	4	0,025	0,4
1,6	2,5	0,016	0,25

In the description number T10150 Revision 1 further characteristics are described.

Valid until 17 June 2018

Description and documentation The gasmeter is described in the description number T10150 Revision 1 and documented in the documentation folder T10150-1, appertaining to this EC-type examination certificate.

Remark This version replaces the former version, with the exception of the documentation folder.

Dordrecht, 26 February 2009  
NMI Certin B.V.

Ing. C. Oosterman  
manager Product Certification

**NMI Certin BV**  
Hugo de Grootplein 1  
3314 EG Dordrecht  
PO Box 394  
3300 AJ Dordrecht, NL  
T +31 78 6332332  
F +31 78 6332309  
certin@nmi.nl  
www.nmi.nl

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see "Regulation objection and appeal against decisions of NMI")

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Reproduction of the complete document is permitted.

